

# Calibration Certification

**Product:** LG01 Series

**Description:** Liquid Flow Switch and Bubble Detector.

The above mentioned products are calibrated to meet the specifications according to the corresponding Sensirion data sheet. Each device is individually tested after its calibration.

Sensirion uses transfer standards for the calibration. These transfer standards are themselves subject to a scheduled calibration procedure. The calibration of the reference itself used for the calibration of the transfer standards is performed by an ISO/IEC 17025 accredited laboratory.

The accreditation body is full member of the International Laboratory Accreditation Cooperation ([www.ilac.org](http://www.ilac.org)). Calibration certificates issued by facilities accredited by a signatory to the ILAC Mutual Recognition Arrangement (MRA) are accepted by all signatories to the ILAC MRA.

This provides traceability of measurement to recognized national standards and to units of measurement realized at the “National Physical Laboratory” (NPL) or other recognized national standards laboratories like “Physikalisch-Technische Bundesanstalt” (PTB) or “National Institute of Standards and Technology” (NIST).

Staefa, October 2013



Dr. Felix Mayer  
Co-CEO, Sensirion AG



Stephan Weber  
Quality Manager, Sensirion AG