

# Calibration Certification

**Product:** LPG 10-xxx Series  
**Description:** Liquid Flow Sensors

The above mentioned products are calibrated to meet the specifications according to the corresponding Sensirion data sheet. Each device is individually tested after its calibration.

Sensirion uses transfer standards for the calibration. These transfer standards are themselves subject to a scheduled calibration procedure. The calibration of the primary reference itself used for the calibration of the transfer standards is performed by an accredited laboratory.

This provides traceability of measurement to recognized national standards and to units of measurement realized at the “National Physical Laboratory” (NPL) or other recognized national standards laboratories like “Physikalisch-Technische Bundesanstalt” (PTB) or “National Institute of Standards and Technology” (NIST).

This certification is valid until 31.12.2018.

Staefa, Januar 2017



Stephan Weber,  
Head of Quality Management, Sensirion AG