

Sensirion AG
 Laubisruetistrasse 50
 8712 Staefa,
 Switzerland

REACH- and RoHS - Declaration

We declare that **these Liquid Flow Meters and accessories** supplied by Sensirion AG comply with

Liquid Flow Meters and accessories	REACH	RoHS
LD20, LD25	yes	yes
LG16	yes	yes
LPG10	yes	yes
LS32	yes	yes
SLF3x	yes	yes
SLG	yes	yes
SLI	yes	yes
SLQ	yes	yes
SLS	yes	yes
Liquid Flow Switch LG01	yes	yes
Eval-Kit for Liquid Flow Meters	yes	yes
SCC1 Sensor Cable series	yes	yes

- the EU regulation concerning the Registration, Evaluation, Authorisation and Restriction of Chemicals, **REACH** (Reg. 1907/2006 /EC),
 i.e. the following limit values are not exceeded
 - Substances of Very High Concerns (SVHC) 1000 ppm
 The information is based on data provided by our suppliers.
 No attempt has been made to verify this information by third party analysis.
- the EU regulation on the Restriction of Hazardous Substances in electrical and electronic equipment, **RoHS** (Reg. 2011/65/EU and the delegated directive 2015/863/EU)
 i.e. the following limit values are not exceeded
 - Lead (Pb), Mercury (Hg), Hexavalent Chromium (CrVI) 1000 ppm
 - Cadmium (Cd) 100 ppm
 - Polybrominated biphenyls (PBB) and Polybrominated diphenyl ethers (PBDE) 1000 ppm
 - Phthalates: DEHP, BBP, DBP and DIBP 1000 ppm
 - Copper alloy in the plugs of the eval-kits: Compliant with exception 6c of 2011/65/EU.

Disclaimer: All information in this declaration is given to the best of our present knowledge and belief. The information given doesn't imply giving a warranty within the meaning of the warranty law.

Signed for and on behalf of Sensirion AG
 Staefa, 27 January 2025


 Dr. Adrian Ruf
 EHS Manager