

Calibration Certification

Name and address of the manufacturer: Sensirion AG
Laubisruestrasse 50
CH-8712 Switzerland

Description: *Gas Flow Sensors*

- *SFM3xxx*
- *SFM4xxx*
- *SFM6xxx*
- *SFM7xxx*
- *SFT40xx*

The above mentioned products are calibrated to meet the specifications according to the corresponding Sensirion data sheet. Each device is individually tested after its calibration.

Sensirion uses transfer standards for the calibration. These transfer standards are themselves subject to a scheduled calibration procedure. The calibration of the reference itself used for the calibration of the transfer standards is performed in accordance with ISO9001:2015.

The accreditation bodies of our references are full members of the International Laboratory Accreditation Cooperation (www.ilac.org). Calibration certificates issued by facilities accredited by a signatory to the ILAC Mutual Recognition Arrangement (MRA) are accepted by all signatories to the ILAC MRA.

Staefa, 16 January 2023



Stephan Weber,
Director,
Head of Quality Management, Sensirion AG



Volker Born
Manager,
Head of Quality Engineering, Sensirion AG