

SCS Directory

Accreditation number: SCS 0158

International Standard: ISO/IEC 17025:2017
 Swiss standard: SN EN ISO/IEC 17025:2018

Sensirion AG
 Feuchte- und
 Temperatursensoren
 Laubisrütistrasse 26
 8712 Stäfa
 Schweiz

Head: Mr. Markus Klumpp
 Responsible for MS: Mr. Markus Klumpp
 Telephone: +41 44 306 40 00
 E-Mail: markus.klumpp@sensirion.com
 Internet: www.sensirion.com
 Initial accreditation: 19.10.2020
 Current accreditation: 19.10.2020 to 18.10.2025
 Scope of accreditation see: www.sas.admin.ch
 (Accredited bodies)

Scope of accreditation as of 19.10.2020

Calibration laboratory for temperature

Calibration and Measurement Capability (CMC)

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Capability \pm ¹⁾	Remarks
TEMPERATURE				
DIGITAL TEMPERATURE SENSOR				
Fixed-point	-30 °C	In-situ surface temperature measurement by comparison	0.31 K	
Fixed-point	5° C		0.14 K	
Fixed-point	70° C		0.12 K	

* / * / * / * / *